

URGENT RECALL NOTICE

COMPLETE ALL-IN-ONE MOISTUREPLUS CONTACT LENS SOLUTION

AMO Canada Company is immediately recalling all lots of Complete All-In-One MoisturePlus contact lens solution. The U.S. Centers for Disease Control and Prevention (CDC) made data available to AMO Friday May 25, 2007, that it had interviewed 46 patients who had developed Acanthamoeba keratitis (AK) infections reported since January 2005. A total of 39 of these patients were soft contact lens wearers, 21 of whom reported using Complete All-In-One MoisturePlus contact lens solution. The CDC estimates a risk of at least seven times greater for those who used Complete All-In-One MoisturePlus contact lens solution versus those who did not. This recall is being conducted on a voluntary basis, in cooperation with Health Canada.

At this time, we are advising patients to discontinue use of AMO Complete All-In-One MoisturePlus and throw away their current pair of contact lenses and their case.

In its place, we are recommending a disinfectant system called Clear Care and will provide you with a complimentary trial kit.

Any Complete All-In-One MoisturePlus solution purchased in our office can be returned and exchanged for Clear Care up until July 31, 2007.



PRIMARY EYE CARE CENTRE